

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081522\
 Data File : VX030624.D
 Acq On : 15 Aug 2022 13:42
 Operator : JC/MD
 Sample : N4193-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VHBLK001

Quant Time: Aug 16 04:29:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 16 04:27:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	245478	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	227002	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	99190	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	84063	53.325	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	106.660%
7) Chloroethane-d5	1.672	69	65548	58.983	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	117.960%
11) 1,1-Dichloroethene-d2	2.306	63	140098	46.982	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.960%
21) 2-Butanone-d5	4.471	46	167304	114.542	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	114.540%
24) Chloroform-d	5.056	84	179588	55.341	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.680%
26) 1,2-Dichloroethane-d4	5.958	65	131575	59.714	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	119.420%
32) Benzene-d6	5.977	84	350560	53.912	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.820%
36) 1,2-Dichloropropane-d6	7.312	67	115062	54.540	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	109.080%
41) Toluene-d8	8.653	98	316452	51.630	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.260%
43) trans-1,3-Dichloroprop...	8.952	79	52182	49.720	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.440%
47) 2-Hexanone-d5	9.384	63	107544	100.010	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.010%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	150960	54.470	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	108.940%
66) 1,2-Dichlorobenzene-d4	12.323	152	105899	55.044	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	110.080%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081522\
Data File : VX030624.D
Acq On : 15 Aug 2022 13:42
Operator : JC/MD
Sample : N4193-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VHBLK001

Quant Time: Aug 16 04:29:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 16 04:27:08 2022
Response via : Initial Calibration

