

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081522\
 Data File : VX030635.D
 Acq On : 15 Aug 2022 18:55
 Operator : JC/MD
 Sample : N4168-11
 Misc : 8.79g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EX9B8

Quant Time: Aug 16 04:32:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 16 04:27:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	284402	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	249319	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.030	152	115089	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	65152	35.673	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.340%
7) Chloroethane-d5	1.654	69	24234	18.822	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	37.640%#
11) 1,1-Dichloroethene-d2	2.270	63	106301	30.769	ug/L	-0.04
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.540%
21) 2-Butanone-d5	4.507	46	114527	67.678	ug/L	0.04
Spiked Amount	100.000	Range	40 - 130	Recovery	=	67.680%
24) Chloroform-d	5.062	84	132957	35.364	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	70.720%
26) 1,2-Dichloroethane-d4	5.958	65	98737	38.678	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.360%
32) Benzene-d6	5.971	84	270682	37.902	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	75.800%
36) 1,2-Dichloropropane-d6	7.312	67	86617	37.382	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	74.760%
41) Toluene-d8	8.653	98	245035	36.400	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	72.800%#
43) trans-1,3-Dichloroprop...	8.952	79	34797	30.187	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	60.380%
47) 2-Hexanone-d5	9.403	63	72544	61.423	ug/L	0.02
Spiked Amount	100.000	Range	45 - 130	Recovery	=	61.420%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	103479	33.996	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	68.000%
66) 1,2-Dichlorobenzene-d4	12.323	152	82807	37.095	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	74.200%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081522\
Data File : VX030635.D
Acq On : 15 Aug 2022 18:55
Operator : JC/MD
Sample : N4168-11
Misc : 8.79g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EX9B8

Quant Time: Aug 16 04:32:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM080422WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 16 04:27:08 2022
Response via : Initial Calibration

