

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081523\
 Data File : VX036991.D
 Acq On : 15 Aug 2023 16:32
 Operator : JC/MD
 Sample : 03989-03
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Aug 16 05:48:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.544	168	106814	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	193086	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	190008	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95060	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	73031	46.208	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	92.420%		
35) Dibromofluoromethane	5.385	113	59347	44.717	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	89.440%		
50) Toluene-d8	8.647	98	233898	49.698	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.400%		
62) 4-Bromofluorobenzene	11.085	95	98119	55.258	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	110.520%		
Target Compounds						
					Qvalue	
13) Acrolein	2.203	56	68	0.318	ug/l #	14
18) Methyl Acetate	2.733	43	579	0.210	ug/l #	53
20) Methylene Chloride	2.776	84	755	0.518	ug/l #	82
29) Tetrahydrofuran	5.056	42	314	0.458	ug/l #	82
31) Cyclohexane	5.550	56	2150	0.983	ug/l #	1
36) 1,1-Dichloropropene	5.544	75	8735	4.416	ug/l #	50
38) Carbon Tetrachloride	5.550	117	11186	5.319	ug/l #	14
51) 4-Methyl-2-Pentanone	8.641	43	1123	0.527	ug/l #	1
76) 1,2,3-Trichloropropane	11.085	75	50301	22.497	ug/l #	38
81) trans-1,4-Dichloro-2-b...	11.085	75	50301	50.990	ug/l #	5


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081523\
Data File : VX036991.D
Acq On : 15 Aug 2023 16:32
Operator : JC/MD
Sample : 03989-03
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Aug 16 05:48:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX036991.D
Acq: 15 Aug 2023 16:32

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:168 Resp: 106814
Ion Ratio Lower Upper
168 100
99 64.7 56.6 85.0



#13
Acrolein
Concen: 0.318 ug/l
RT: 2.203 min Scan# 183
Delta R.T. -0.036 min
Lab File: VX036991.D
Acq: 15 Aug 2023 16:32


Tgt Ion: 56 Resp: 68
Ion Ratio Lower Upper
56 100
55 0.0 57.1 85.7#








#18
Methyl Acetate
Concen: 0.210 ug/l
RT: 2.733 min Scan# 21
Delta R.T. 0.025 min
Lab File: VX036991.D
Acq: 15 Aug 2023 16:32

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 43 Resp: 579
Ion Ratio Lower Upper
43 100
74 0.0 17.8 26.8#



#20
Methylene Chloride
Concen: 0.518 ug/l
RT: 2.776 min Scan# 277
Delta R.T. -0.012 min
Lab File: VX036991.D
Acq: 15 Aug 2023 16:32


Tgt Ion: 84 Resp: 755
Ion Ratio Lower Upper
84 100
49 104.1 103.6 155.4
51 27.2 33.8 50.6#
86 65.1 49.4 74.0








#29
Tetrahydrofuran
Concen: 0.458 ug/l
RT: 5.056 min Scan# 611
Delta R.T. 0.049 min
Lab File: VX036991.D
Acq: 15 Aug 2023 16:32

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 42 Resp: 314
Ion Ratio Lower Upper
42 100
72 30.6 34.2 51.4#
71 30.3 32.2 48.2#



#31
Cyclohexane
Concen: 0.983 ug/l
RT: 5.550 min Scan# 732
Delta R.T. 0.079 min
Lab File: VX036991.D
Acq: 15 Aug 2023 16:32

Tgt Ion: 56 Resp: 2150
Ion Ratio Lower Upper
56 100
69 194.3 23.4 35.2#
84 138.5 65.9 98.9#









#33
 1,2-Dichloroethane-d4
 Concen: 46.208 ug/l
 RT: 5.958 min Scan# 798
 Delta R.T. 0.006 min
 Lab File: VX036991.D
 Acq: 15 Aug 2023 16:32




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 65 Resp: 73031
 Ion Ratio Lower Upper
 65 100
 67 52.1 0.0 100.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.757 min Scan# 930
 Delta R.T. -0.006 min
 Lab File: VX036991.D
 Acq: 15 Aug 2023 16:32


Tgt Ion: 114 Resp: 193086
 Ion Ratio Lower Upper
 114 100
 63 20.7 0.0 45.0
 88 16.4 0.0 36.6








#35
Dibromofluoromethane
Concen: 44.717 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX036991.D
Acq: 15 Aug 2023 16:32

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:113 Resp: 59347
Ion Ratio Lower Upper
113 100
111 101.8 82.9 124.3
192 18.6 13.3 19.9



#36
1,1-Dichloropropene
Concen: 4.416 ug/l
RT: 5.544 min Scan# 731
Delta R.T. -0.152 min
Lab File: VX036991.D
Acq: 15 Aug 2023 16:32


Tgt Ion: 75 Resp: 8735
Ion Ratio Lower Upper
75 100
110 8.8 17.0 50.9#
77 0.0 24.7 37.1#








#38
Carbon Tetrachloride
Concen: 5.319 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.128 min
Lab File: VX036991.D
Acq: 15 Aug 2023 16:32

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:117 Resp: 11186
Ion Ratio Lower Upper
117 100
119 4.1 78.2 117.2#
121 0.0 24.4 36.6#



#50
Toluene-d8
Concen: 49.698 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX036991.D
Acq: 15 Aug 2023 16:32

Tgt Ion: 98 Resp: 233898
Ion Ratio Lower Upper
98 100
100 66.6 53.1 79.7





#51

4-Methyl-2-Pentanone

Concen: 0.527 ug/l

RT: 8.641 min Scan# 1123

Delta R.T. 0.067 min

Lab File: VX036991.D

Acq: 15 Aug 2023 16:32

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 1123

Ion Ratio Lower Upper

43 100

58 171.2 30.4 45.6#



#62

4-Bromofluorobenzene

Concen: 55.258 ug/l

RT: 11.085 min Scan# 1640

Delta R.T. 0.006 min

Lab File: VX036991.D

Acq: 15 Aug 2023 16:32




Tgt Ion: 95 Resp: 98119


Ion Ratio Lower Upper

95 100

174 71.0 0.0 135.6

176 70.4 0.0 131.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VX036991.D

Acq: 15 Aug 2023 16:32

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:117 Resp: 190008

Ion Ratio Lower Upper

117 100

82 55.9 46.2 69.4

119 31.7 25.5 38.3



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.000 min

Lab File: VX036991.D

Acq: 15 Aug 2023 16:32




Tgt Ion:152 Resp: 95060


Ion Ratio Lower Upper

152 100

115 62.6 43.3 129.9

150 156.6 0.0 345.0





#76

1,2,3-Trichloropropane

Concen: 22.497 ug/l

RT: 11.085 min Scan# 1

Delta R.T. -0.152 min

Lab File: VX036991.D

Acq: 15 Aug 2023 16:32

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 75 Resp: 50301

Ion Ratio Lower Upper

75 100

77 1.0 19.1 57.5#



#81

trans-1,4-Dichloro-2-butene

Concen: 50.990 ug/l

RT: 11.085 min Scan# 1640

Delta R.T. 0.067 min

Lab File: VX036991.D

Acq: 15 Aug 2023 16:32

Tgt Ion: 75 Resp: 50301

Ion Ratio Lower Upper

75 100

53 0.0 90.5 135.7#

89 0.0 36.9 55.3#

