

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081523\
 Data File : VX036979.D
 Acq On : 15 Aug 2023 11:30
 Operator : JC/MD
 Sample : VX0815MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0815MBL01

Quant Time: Aug 15 16:52:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	127561	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	234181	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231260	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	109825	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	89461	47.398	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.800%
35) Dibromofluoromethane	5.385	113	75934	47.175	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.340%
50) Toluene-d8	8.646	98	284227	49.794	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.580%
62) 4-Bromofluorobenzene	11.079	95	115521	53.642	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081523\
Data File : VX036979.D
Acq On : 15 Aug 2023 11:30
Operator : JC/MD
Sample : VX0815MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0815MBL01

Quant Time: Aug 15 16:52:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

