

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081523\
 Data File : VX036985.D
 Acq On : 15 Aug 2023 14:07
 Operator : JC/MD
 Sample : 03997-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-119S1-20230811

Quant Time: Aug 16 05:33:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	122513	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	225621	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208938	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91087	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86525	47.731	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	95.460%		
35) Dibromofluoromethane	5.391	113	71779	46.285	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.580%		
50) Toluene-d8	8.647	98	269106	48.933	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.860%		
62) 4-Bromofluorobenzene	11.079	95	102324	49.316	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	98.640%		
Target Compounds						Qvalue
44) Trichloroethene	7.129	130	8329	4.738	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081523\
Data File : VX036985.D
Acq On : 15 Aug 2023 14:07
Operator : JC/MD
Sample : 03997-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-119S1-20230811

Quant Time: Aug 16 05:33:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

