

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081523\
 Data File : VX036989.D
 Acq On : 15 Aug 2023 15:44
 Operator : JC/MD
 Sample : 03989-01
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B-PP25

Quant Time: Aug 15 16:53:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	103510	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	194749	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	194084	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100997	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	73984	48.306	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.620%
35) Dibromofluoromethane	5.385	113	61284	45.782	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.560%
50) Toluene-d8	8.647	98	236042	49.725	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.440%
62) 4-Bromofluorobenzene	11.079	95	101745	56.811	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081523\
Data File : VX036989.D
Acq On : 15 Aug 2023 15:44
Operator : JC/MD
Sample : 03989-01
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B-PP25

Quant Time: Aug 15 16:53:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

