

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081523\
 Data File : VX036994.D
 Acq On : 15 Aug 2023 17:44
 Operator : JC/MD
 Sample : 03899-06DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 15017DL

Quant Time: Aug 16 05:34:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	139363	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	262252	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	256349	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	122511	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	99624	48.312	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.620%		
35) Dibromofluoromethane	5.385	113	82112	45.553	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.100%		
50) Toluene-d8	8.646	98	316948	49.583	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.160%		
62) 4-Bromofluorobenzene	11.079	95	129551	53.717	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	107.440%		
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.965	59	37013	75.817	ug/l #	92
16) Acetone	2.379	43	467329	546.146	ug/l	92
20) Methylene Chloride	2.788	84	3065	1.611	ug/l #	80
25) 2-Butanone	4.562	43	20240	14.579	ug/l	92
43) Isopropyl Acetate	6.348	43	13585	2.595	ug/l	99
48) Methyl methacrylate	7.695	41	16856	6.924	ug/l #	85
51) 4-Methyl-2-Pentanone	8.573	43	54941	18.966	ug/l	99
52) Toluene	8.720	92	234692	47.916	ug/l	99
64) Tetrachloroethene	9.274	164	47943	25.074	ug/l	98
69) o-Xylene	10.640	106	3968	1.032	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081523\
Data File : VX036994.D
Acq On : 15 Aug 2023 17:44
Operator : JC/MD
Sample : 03899-06DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
15017DL

Quant Time: Aug 16 05:34:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

