

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081524\
 Data File : VX043053.D
 Acq On : 15 Aug 2024 11:18
 Operator : JC/MD
 Sample : P3410-01DL 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-S-864-9-14-073024DL

Quant Time: Aug 16 02:41:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081424W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 02:11:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	91818	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	157091	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	137603	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60840	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	54956	45.550	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.100%
35) Dibromofluoromethane	5.391	113	48646	50.056	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.120%
50) Toluene-d8	8.647	98	189790	50.591	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.180%
62) 4-Bromofluorobenzene	11.079	95	63902	46.977	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	93.960%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.319	101	686	0.665	ug/l	89
27) cis-1,2-Dichloroethene	4.483	96	68434	49.087	ug/l	82
44) Trichloroethene	7.129	130	12927	12.411	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081524\
Data File : VX043053.D
Acq On : 15 Aug 2024 11:18
Operator : JC/MD
Sample : P3410-01DL 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-S-864-9-14-073024DL

Quant Time: Aug 16 02:41:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081424W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 02:11:44 2024
Response via : Initial Calibration

