

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081524\
 Data File : VX043054.D
 Acq On : 15 Aug 2024 11:41
 Operator : JC/MD
 Sample : P3542-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-SOIL-REF-6

Quant Time: Aug 16 01:39:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081424W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 04:22:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	91276	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	160387	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	141333	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62935	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	54021	45.041	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.080%
35) Dibromofluoromethane	5.385	113	46789	47.156	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.320%
50) Toluene-d8	8.647	98	183351	47.870	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.740%
62) 4-Bromofluorobenzene	11.079	95	62399	44.930	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	89.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081524\
Data File : VX043054.D
Acq On : 15 Aug 2024 11:41
Operator : JC/MD
Sample : P3542-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-SOIL-REF-6

Quant Time: Aug 16 01:39:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081424W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 04:22:11 2024
Response via : Initial Calibration

