

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081619\
 Data File : VX011572.D
 Acq On : 16 Aug 2019 11:32
 Operator : JC/SP
 Sample : VX0816MBL01
 Misc : 5.0g/10mL/100uL/10mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0816MBL01

Quant Time: Aug 19 01:26:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	445307	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	616295	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	530368	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	256050	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	211192	51.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.28%	
35) Dibromofluoromethane	5.49	113	187111	46.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.84%	
50) Toluene-d8	8.71	98	693525	55.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.44%	
62) 4-Bromofluorobenzene	11.13	95	237389	44.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.56%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081619\
Data File : VX011572.D
Acq On : 16 Aug 2019 11:32
Operator : JC/SP
Sample : VX0816MBL01
Misc : 5.0g/10mL/100uL/10mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0816MBL01

Quant Time: Aug 19 01:26:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

