

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081619\
 Data File : VX011578.D
 Acq On : 16 Aug 2019 14:03
 Operator : JC/SP
 Sample : K4279-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-ER-2019-1

Quant Time: Aug 19 01:52:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	449855	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	614818	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	548461	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	255605	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	214807	51.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.98%	
35) Dibromofluoromethane	5.49	113	186293	46.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.66%	
50) Toluene-d8	8.71	98	695745	55.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.06%	
62) 4-Bromofluorobenzene	11.13	95	237183	44.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.70%	
Target Compounds						
16) Acetone	2.43	43	14131	6.737	ug/l	93
20) Methylene Chloride	2.85	84	34355	7.800	ug/l	97
29) Tetrahydrofuran	5.12	42	84465	39.100	ug/l	98
67) Ethyl Benzene	10.25	91	8143	0.417	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081619\
Data File : VX011578.D
Acq On : 16 Aug 2019 14:03
Operator : JC/SP
Sample : K4279-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-ER-2019-1

Quant Time: Aug 19 01:52:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

