

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081619\
 Data File : VX011586.D
 Acq On : 16 Aug 2019 17:11
 Operator : JC/SP
 Sample : K4223-12
 Misc : 4.85g/5mL/100uL/10mL/MSVOA_X/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP-6

Quant Time: Aug 19 03:26:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	434727	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	598249	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	538252	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	292799	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	203108	50.38	ug/l	0.00
Spiked Amount			Recovery	=	100.76%	
35) Dibromofluoromethane	5.49	113	168314	43.01	ug/l	0.00
Spiked Amount			Recovery	=	86.02%	
50) Toluene-d8	8.71	98	682812	56.00	ug/l	0.00
Spiked Amount			Recovery	=	112.00%	
62) 4-Bromofluorobenzene	11.13	95	251562	48.88	ug/l	0.00
Spiked Amount			Recovery	=	97.76%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.47	43	13521	6.670	ug/l	90
18) Methyl Acetate	2.78	43	114634	18.234	ug/l	96
31) Cyclohexane	5.56	56	10009	1.674	ug/l	86
39) Methylcyclohexane	7.45	83	7573	1.199	ug/l	98
67) Ethyl Benzene	10.25	91	33721	1.760	ug/l	95
68) m/p-Xylenes	10.35	106	49302	6.704	ug/l	96
84) 1,2,4-Trimethylbenzene	11.80	105	18200	1.136	ug/l	97
95) Naphthalene	13.83	128	21587	1.051	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081619\
Data File : VX011586.D
Acq On : 16 Aug 2019 17:11
Operator : JC/SP
Sample : K4223-12
Misc : 4.85g/5mL/100uL/10mL/MSVOA_X/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP-6

Quant Time: Aug 19 03:26:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

