

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081619\
 Data File : VX011589.D
 Acq On : 16 Aug 2019 18:22
 Operator : JC/SP
 Sample : K4397-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Aug 19 03:34:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	486010	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	659186	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	588470	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	318506	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	221473	49.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.28%	
35) Dibromofluoromethane	5.49	113	194535	45.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.24%	
50) Toluene-d8	8.71	98	750246	55.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.70%	
62) 4-Bromofluorobenzene	11.13	95	276938	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
Target Compounds						
					Qvalue	
16) Acetone	2.44	43	6764	2.985	ug/l #	85
31) Cyclohexane	5.58	56	2279	0.341	ug/l #	61
39) Methylcyclohexane	7.46	83	4498	0.646	ug/l	98
52) Toluene	8.78	92	3992	0.358	ug/l	84
67) Ethyl Benzene	10.25	91	8542	0.408	ug/l #	87
68) m/p-Xylenes	10.35	106	11806	1.468	ug/l	97
84) 1,2,4-Trimethylbenzene	11.80	105	11545	0.662	ug/l	93
95) Naphthalene	13.83	128	11159	0.499	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081619\
Data File : VX011589.D
Acq On : 16 Aug 2019 18:22
Operator : JC/SP
Sample : K4397-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Aug 19 03:34:33 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

