

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081619\
 Data File : VX011591.D
 Acq On : 16 Aug 2019 19:09
 Operator : JC/SP
 Sample : K4397-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Aug 19 03:39:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	480408	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	661159	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	586965	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	299954	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	224177	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
35) Dibromofluoromethane	5.49	113	193139	44.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.32%	
50) Toluene-d8	8.71	98	749231	55.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.20%	
62) 4-Bromofluorobenzene	11.13	95	267633	47.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.12%	
Target Compounds						
16) Acetone	2.43	43	6213	2.773	ug/l	91
39) Methylcyclohexane	7.46	83	2405	0.345	ug/l	89
68) m/p-Xylenes	10.35	106	7039	0.878	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081619\
Data File : VX011591.D
Acq On : 16 Aug 2019 19:09
Operator : JC/SP
Sample : K4397-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Aug 19 03:39:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

