

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081619\
 Data File : VX011593.D
 Acq On : 16 Aug 2019 19:56
 Operator : JC/SP
 Sample : K4368-01RE
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PH2-0134RE

Quant Time: Aug 19 03:43:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	454141	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	622295	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	551638	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	270566	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	212977	50.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.14%	
35) Dibromofluoromethane	5.49	113	182724	44.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.78%	
50) Toluene-d8	8.71	98	705314	55.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.22%	
62) 4-Bromofluorobenzene	11.13	95	252501	47.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.34%	
Target Compounds						
					Qvalue	
31) Cyclohexane	5.57	56	9076	1.453	ug/l	95
39) Methylcyclohexane	7.46	83	18348	2.792	ug/l #	77
67) Ethyl Benzene	10.25	91	942331	47.989	ug/l	98
73) Isopropylbenzene	11.02	105	207016	11.791	ug/l	100
78) n-propylbenzene	11.36	91	120558	6.368	ug/l	100
84) 1,2,4-Trimethylbenzene	11.80	105	7392	0.499	ug/l	97
85) sec-Butylbenzene	11.94	105	15613	0.921	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081619\
Data File : VX011593.D
Acq On : 16 Aug 2019 19:56
Operator : JC/SP
Sample : K4368-01RE
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PH2-0134RE

Quant Time: Aug 19 03:43:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

