

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081619\
 Data File : VX011597.D
 Acq On : 16 Aug 2019 22:40
 Operator : JC/SP
 Sample : VX0816WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0816WBL02

Quant Time: Aug 19 03:49:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	459662	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	636470	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	561701	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	274884	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	216814	50.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.72%	
35) Dibromofluoromethane	5.49	113	187476	45.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.06%	
50) Toluene-d8	8.71	98	715153	55.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.26%	
62) 4-Bromofluorobenzene	11.13	95	251405	45.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.84%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081619\
Data File : VX011597.D
Acq On : 16 Aug 2019 22:40
Operator : JC/SP
Sample : VX0816WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0816WBL02

Quant Time: Aug 19 03:49:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

