

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081619\
 Data File : VX011609.D
 Acq On : 17 Aug 2019 03:21
 Operator : JC/SP
 Sample : K4265-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 REACTOR-4-A4-(0-1)

Quant Time: Aug 19 04:37:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	447014	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	626390	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	523613	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	268630	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	212638	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
35) Dibromofluoromethane	5.49	113	183682	44.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.66%	
50) Toluene-d8	8.71	98	695326	54.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.94%	
62) 4-Bromofluorobenzene	11.13	95	247697	45.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.94%	
Target Compounds						
16) Acetone	2.44	43	13215	6.340	ug/l	96
43) Isopropyl Acetate	6.44	43	77430	8.054	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081619\
Data File : VX011609.D
Acq On : 17 Aug 2019 03:21
Operator : JC/SP
Sample : K4265-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
REACTOR-4-A4-(0-1)

Quant Time: Aug 19 04:37:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

