

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081619\
 Data File : VX011610.D
 Acq On : 17 Aug 2019 03:45
 Operator : JC/SP
 Sample : K4265-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 REACTOR-4-A1-(3.4)

Quant Time: Aug 19 04:43:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	451596	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	621188	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	557016	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	276201	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	212742	50.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.60%	
35) Dibromofluoromethane	5.49	113	186249	45.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.68%	
50) Toluene-d8	8.71	98	710149	56.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.20%	
62) 4-Bromofluorobenzene	11.13	95	249952	46.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.56%	
Target Compounds						
16) Acetone	2.43	43	11538	5.479	ug/l	99
20) Methylene Chloride	2.85	84	9794	2.215	ug/l #	95
43) Isopropyl Acetate	6.44	43	80581	8.452	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081619\
Data File : VX011610.D
Acq On : 17 Aug 2019 03:45
Operator : JC/SP
Sample : K4265-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
REACTOR-4-A1-(3.4)

Quant Time: Aug 19 04:43:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

