

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081619\
 Data File : VX011611.D
 Acq On : 17 Aug 2019 04:08
 Operator : JC/SP
 Sample : K4265-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 REACTOR-4-A2-(3.4)

Quant Time: Aug 19 04:45:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	449143	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	625397	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	545896	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	267437	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	213312	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
35) Dibromofluoromethane	5.49	113	182702	44.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.34%	
50) Toluene-d8	8.71	98	708794	55.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.22%	
62) 4-Bromofluorobenzene	11.13	95	245239	45.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.18%	
Target Compounds						
16) Acetone	2.43	43	12457	5.948	ug/l	99
20) Methylene Chloride	2.85	84	8838	2.010	ug/l	99
43) Isopropyl Acetate	6.44	43	77266	8.050	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081619\
Data File : VX011611.D
Acq On : 17 Aug 2019 04:08
Operator : JC/SP
Sample : K4265-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
REACTOR-4-A2-(3.4)

Quant Time: Aug 19 04:45:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

