

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081619\
 Data File : VX011613.D
 Acq On : 17 Aug 2019 04:55
 Operator : JC/SP
 Sample : K4265-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 REACTOR-4-A4-(3.4)

Quant Time: Aug 19 04:48:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	446222	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	613935	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	557039	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	275927	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	208400	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
35) Dibromofluoromethane	5.49	113	182008	45.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.66%	
50) Toluene-d8	8.71	98	702069	56.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.22%	
62) 4-Bromofluorobenzene	11.13	95	250873	47.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.00%	
Target Compounds						
16) Acetone	2.43	43	10663	5.125	ug/l	99
20) Methylene Chloride	2.85	84	10171	2.328	ug/l	94
43) Isopropyl Acetate	6.44	43	81053	8.602	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081619\
Data File : VX011613.D
Acq On : 17 Aug 2019 04:55
Operator : JC/SP
Sample : K4265-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
REACTOR-4-A4-(3.4)

Quant Time: Aug 19 04:48:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

