

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081619\
 Data File : VX011615.D
 Acq On : 17 Aug 2019 05:42
 Operator : JC/SP
 Sample : K4265-19
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 R1-R2-(4.3)

Quant Time: Aug 19 04:52:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	442049	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	608523	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	549017	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	270416	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	208079	50.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.52%	
35) Dibromofluoromethane	5.49	113	179415	45.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.16%	
50) Toluene-d8	8.71	98	699565	56.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.82%	
62) 4-Bromofluorobenzene	11.13	95	248056	47.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.78%	
Target Compounds						
16) Acetone	2.43	43	13117	6.364	ug/l	94
20) Methylene Chloride	2.85	84	4580	1.058	ug/l	94
43) Isopropyl Acetate	6.44	43	74155	7.940	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081619\
Data File : VX011615.D
Acq On : 17 Aug 2019 05:42
Operator : JC/SP
Sample : K4265-19
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
R1-R2-(4.3)

Quant Time: Aug 19 04:52:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration

