

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081619\
 Data File : VX011572.D
 Acq On : 16 Aug 2019 11:32
 Operator : JC/SP
 Sample : VX0816MBL01
 Misc : 5.0g/10mL/100uL/10mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0816MBL01

Quant Time: Aug 19 01:26:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	445307	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	616295	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	530368	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	256050	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	211192	51.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.28%	
35) Dibromofluoromethane	5.49	113	187111	46.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.84%	
50) Toluene-d8	8.71	98	693525	55.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.44%	
62) 4-Bromofluorobenzene	11.13	95	237389	44.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.56%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081619\
Data File : VX011572.D
Acq On : 16 Aug 2019 11:32
Operator : JC/SP
Sample : VX0816MBL01
Misc : 5.0g/10mL/100uL/10mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0816MBL01

Quant Time: Aug 19 01:26:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 738

Delta R.T. 0.00 min

Lab File: VX011572.D

Acq: 16 Aug 2019 11:32

Instrument :

MSVOA_X

ClientSampled :


VX0816MBL01

Tot Ion:168 Resp: 445307

Ion Ratio Lower Upper

168 100

99 47.0 37.3 55.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.65



150000

100000

50000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 51.141 ug/l

RT: 6.05 min Scan# 803

Delta R.T. -0.01 min

Lab File: VX011572.D


Acq: 16 Aug 2019 11:32

Tgt Ion: 65 Resp: 211192

Ion Ratio Lower Upper

65 100

67 52.5 0.0 107.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.05

80000


60000


40000

20000

0

Time-->





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 934

Delta R.T. -0.01 min

Lab File: VX011572.D

Acq: 16 Aug 2019 11:32

Instrument :

MSVOA_X

ClientSampleId :

VX0816MBL01


Tot Ion:114 Resp: 616295

Ion Ratio Lower Upper

114 100

63 18.1 0.0 37.8




88 14.3 0.0 30.2



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 46.419 ug/l

RT: 5.49 min Scan# 711

Delta R.T. -0.01 min

Lab File: VX011572.D

Acq: 16 Aug 2019 11:32


Tgt Ion:113 Resp: 187111

Ion Ratio Lower Upper

113 100

111 101.1 81.1 121.7



192 25.5 20.8 31.2




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 55.217 ug/l

RT: 8.71 min Scan# 1239

Delta R.T. -0.01 min

Lab File: VX011572.D

Acq: 16 Aug 2019 11:32

Instrument :

MSVOA_X

ClientSampleId :





VX0816MBL01

Tot Ion: 98 Resp: 693525

Ion Ratio Lower Upper

98 100

100 64.9 52.4 78.6



#62

4-Bromofluorobenzene

Concen: 44.779 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX011572.D

Acq: 16 Aug 2019 11:32




Tgt Ion: 95 Resp: 237389


Ion Ratio Lower Upper

95 100

174 98.4 0.0 200.6

176 96.5 0.0 196.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. 0.00 min
Lab File: VX011572.D
Acq: 16 Aug 2019 11:32

Instrument :
MSVOA_X
ClientSampleId :
VX0816MBL01

Tot Ion	Ratio	Lower	Upper
117	100		
82	51.3	42.0	63.0
119	32.3	26.5	39.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.07 min Scan# 1791
Delta R.T. 0.00 min
Lab File: VX011572.D
Acq: 16 Aug 2019 11:32

Tgt Ion	Ratio	Lower	Upper
152	100		
115	53.9	37.3	111.9
150	155.8	0.0	346.2

