

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081619\
 Data File : VX011597.D
 Acq On : 16 Aug 2019 22:40
 Operator : JC/SP
 Sample : VX0816WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0816WBL02

Quant Time: Aug 19 03:49:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 16 04:15:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	459662	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	636470	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	561701	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	274884	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	216814	50.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.72%	
35) Dibromofluoromethane	5.49	113	187476	45.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.06%	
50) Toluene-d8	8.71	98	715153	55.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.26%	
62) 4-Bromofluorobenzene	11.13	95	251405	45.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.84%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081619\
Data File : VX011597.D
Acq On : 16 Aug 2019 22:40
Operator : JC/SP
Sample : VX0816WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0816WBL02

Quant Time: Aug 19 03:49:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081519W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 16 04:15:26 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 738

Delta R.T. 0.00 min

Lab File: VX011597.D

Acq: 16 Aug 2019 22:40

Instrument :

MSVOA_X

ClientSampleId :


VX0816WBL02

Tot Ion:168 Resp: 459662

Ion Ratio Lower Upper




168 100

99 45.8 37.3 55.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 50.863 ug/l

RT: 6.06 min Scan# 804

Delta R.T. 0.00 min

Lab File: VX011597.D


Acq: 16 Aug 2019 22:40

Tgt Ion: 65 Resp: 216814

Ion Ratio Lower Upper



65 100


67 53.1 0.0 107.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.85 min Scan# 935

Delta R.T. 0.00 min

Lab File: VX011597.D

Acq: 16 Aug 2019 22:40

Instrument :

MSVOA_X

ClientSampleId :

VX0816WBL02


Tot Ion:114 Resp: 636470

Ion Ratio Lower Upper

114 100

63 18.3 0.0 37.8




88 14.6 0.0 30.2



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 45.035 ug/l

RT: 5.49 min Scan# 712

Delta R.T. 0.00 min

Lab File: VX011597.D

Acq: 16 Aug 2019 22:40


Tgt Ion:113 Resp: 187476

Ion Ratio Lower Upper

113 100

111 103.3 81.1 121.7



192 25.7 20.8 31.2




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 55.134 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX011597.D

Acq: 16 Aug 2019 22:40

Instrument :

MSVOA_X

ClientSampleId :





VX0816WBL02

Tot Ion: 98 Resp: 715153

Ion Ratio Lower Upper

98 100

100 64.7 52.4 78.6



#62

4-Bromofluorobenzene

Concen: 45.920 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX011597.D

Acq: 16 Aug 2019 22:40




Tgt Ion: 95 Resp: 251405


Ion Ratio Lower Upper

95 100

174 98.9 0.0 200.6

176 96.6 0.0 196.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. 0.00 min
Lab File: VX011597.D
Acq: 16 Aug 2019 22:40

Instrument :
MSVOA_X
ClientSampleId :
VX0816WBL02

Tot Ion:117 Resp: 561701
Ion Ratio Lower Upper
117 100
82 52.2 42.0 63.0
119 31.8 26.5 39.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.01 min
Lab File: VX011597.D
Acq: 16 Aug 2019 22:40

Tgt Ion:152 Resp: 274884
Ion Ratio Lower Upper
152 100
115 52.4 37.3 111.9
150 154.5 0.0 346.2

