

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
 Data File : VX023735.D
 Acq On : 16 Aug 2021 14:03
 Operator : JC/MD
 Sample : M3371-02RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JET-GROUTRE

Quant Time: Aug 17 06:16:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	158658	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	271188	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	250484	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103492	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	111010	56.091	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.180%
35) Dibromofluoromethane	5.397	113	43795	25.829	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	51.660%#
50) Toluene-d8	8.653	98	331375	53.193	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.380%
62) 4-Bromofluorobenzene	11.085	95	114258	53.938	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.880%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	116639	120.142	ug/l	92
20) Methylene Chloride	2.794	84	13308	7.047	ug/l #	81
25) 2-Butanone	4.574	43	8684	5.762	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
Data File : VX023735.D
Acq On : 16 Aug 2021 14:03
Operator : JC/MD
Sample : M3371-02RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JET-GROUTRE

Quant Time: Aug 17 06:16:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

