

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
 Data File : VX023738.D
 Acq On : 16 Aug 2021 15:13
 Operator : JC/MD
 Sample : M3398-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SWCS-04-0510

Quant Time: Aug 17 06:17:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	145156	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	253037	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	237932	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96804	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103928	57.397	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.800%
35) Dibromofluoromethane	5.397	113	81136	51.284	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.560%
50) Toluene-d8	8.653	98	313380	53.913	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.820%
62) 4-Bromofluorobenzene	11.085	95	105592	53.422	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.840%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	63253	71.213	ug/l	91
20) Methylene Chloride	2.794	84	14286	8.269	ug/l	89
25) 2-Butanone	4.574	43	15267	11.072	ug/l	93
43) Isopropyl Acetate	6.355	43	21839	4.883	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
Data File : VX023738.D
Acq On : 16 Aug 2021 15:13
Operator : JC/MD
Sample : M3398-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SWCS-04-0510

Quant Time: Aug 17 06:17:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

