

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
 Data File : VX023741.D
 Acq On : 16 Aug 2021 16:24
 Operator : JC/MD
 Sample : M3407-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 26N

Quant Time: Aug 17 06:18:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	143132	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	245828	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236919	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97744	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	101280	56.725	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.460%
35) Dibromofluoromethane	5.397	113	80224	52.194	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.380%
50) Toluene-d8	8.653	98	306656	54.304	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.600%
62) 4-Bromofluorobenzene	11.085	95	106680	55.556	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.120%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	17999	20.551	ug/l	96
20) Methylene Chloride	2.794	84	12715	7.463	ug/l #	93
43) Isopropyl Acetate	6.355	43	19667	4.527	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
Data File : VX023741.D
Acq On : 16 Aug 2021 16:24
Operator : JC/MD
Sample : M3407-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
26N

Quant Time: Aug 17 06:18:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

