

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
 Data File : VX023742.D
 Acq On : 16 Aug 2021 16:48
 Operator : JC/MD
 Sample : M3407-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 26S

Quant Time: Aug 17 06:18:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	152360	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	259081	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	251290	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103415	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107774	56.707	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.420%
35) Dibromofluoromethane	5.397	113	84753	52.320	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.640%
50) Toluene-d8	8.653	98	326513	54.862	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	109.720%
62) 4-Bromofluorobenzene	11.085	95	112704	55.690	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.380%
Target Compounds						
						Qvalue
16) Acetone	2.385	43	15423	16.543	ug/l	# 86
20) Methylene Chloride	2.794	84	11767	6.489	ug/l	92
43) Isopropyl Acetate	6.348	43	23616	5.158	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
Data File : VX023742.D
Acq On : 16 Aug 2021 16:48
Operator : JC/MD
Sample : M3407-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
26S

Quant Time: Aug 17 06:18:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

