

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
 Data File : VX023743.D
 Acq On : 16 Aug 2021 17:11
 Operator : JC/MD
 Sample : M3407-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 27N

Quant Time: Aug 17 06:18:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	137161	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	236605	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	220332	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91094	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100605	58.800	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 117.600%#	
35) Dibromofluoromethane	5.397	113	76764	51.890	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.780%	
50) Toluene-d8	8.652	98	294666	54.214	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.420%	
62) 4-Bromofluorobenzene	11.085	95	100120	54.172	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 108.340%	
Target Compounds						
						Qvalue
16) Acetone	2.385	43	21187	25.244	ug/l	90
20) Methylene Chloride	2.794	84	11072	6.782	ug/l	92
43) Isopropyl Acetate	6.348	43	18763	4.487	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
Data File : VX023743.D
Acq On : 16 Aug 2021 17:11
Operator : JC/MD
Sample : M3407-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
27N

Quant Time: Aug 17 06:18:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

