

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX081621\  
 Data File : VX023744.D  
 Acq On : 16 Aug 2021 17:35  
 Operator : JC/MD  
 Sample : M3407-18  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 27S

Quant Time: Aug 17 06:18:53 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072221W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jul 23 06:05:38 2021  
 Response via : Initial Calibration

| Compound                    | R.T. QIon |       | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|-----------|-------|----------|----------|------------|----------|
| -----                       |           |       |          |          |            |          |
| Internal Standards          |           |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.562     | 168   | 136940   | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.769     | 114   | 235181   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.055    | 117   | 220847   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024    | 152   | 90346    | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |           |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.964     | 65    | 99202    | 58.074   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 78 - 117 | Recovery | = 116.140% |          |
| 35) Dibromofluoromethane    | 5.397     | 113   | 76964    | 52.340   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 75 - 124 | Recovery | = 104.680% |          |
| 50) Toluene-d8              | 8.653     | 98    | 295099   | 54.623   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 92 - 112 | Recovery | = 109.240% |          |
| 62) 4-Bromofluorobenzene    | 11.085    | 95    | 96946    | 52.772   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 83 - 123 | Recovery | = 105.540% |          |
| Target Compounds            |           |       |          |          |            |          |
|                             |           |       |          |          |            | Qvalue   |
| 16) Acetone                 | 2.386     | 43    | 16625    | 19.840   | ug/l       | 93       |
| 20) Methylene Chloride      | 2.794     | 84    | 10517    | 6.452    | ug/l #     | 82       |
| 43) Isopropyl Acetate       | 6.348     | 43    | 21717    | 5.225    | ug/l       | 92       |
| -----                       |           |       |          |          |            |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX081621\  
Data File : VX023744.D  
Acq On : 16 Aug 2021 17:35  
Operator : JC/MD  
Sample : M3407-18  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
27S

Quant Time: Aug 17 06:18:53 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072221W.M  
Quant Title : SW846 8260  
QLast Update : Fri Jul 23 06:05:38 2021  
Response via : Initial Calibration

