

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
 Data File : VX023745.D
 Acq On : 16 Aug 2021 17:58
 Operator : JC/MD
 Sample : M3407-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 28N

Quant Time: Aug 17 06:19:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	140308	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	243573	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	232454	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95431	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102185	58.384	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 116.760%	
35) Dibromofluoromethane	5.397	113	79183	51.994	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.980%	
50) Toluene-d8	8.653	98	304444	54.411	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.820%	
62) 4-Bromofluorobenzene	11.085	95	104882	55.125	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 110.260%	
Target Compounds						Qvalue
16) Acetone	2.386	43	22287	25.959	ug/l	89
20) Methylene Chloride	2.794	84	11620	6.958	ug/l #	90
43) Isopropyl Acetate	6.355	43	12824	2.979	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
Data File : VX023745.D
Acq On : 16 Aug 2021 17:58
Operator : JC/MD
Sample : M3407-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
28N

Quant Time: Aug 17 06:19:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

