

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
 Data File : VX023746.D
 Acq On : 16 Aug 2021 18:22
 Operator : JC/MD
 Sample : M3407-24
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 28S

Quant Time: Aug 17 06:19:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	145816	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	251633	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	236346	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91890	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103848	57.093	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 114.180%	
35) Dibromofluoromethane	5.397	113	81691	51.923	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.840%	
50) Toluene-d8	8.653	98	313988	54.319	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.640%	
62) 4-Bromofluorobenzene	11.085	95	104117	52.970	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.940%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	16514	18.508	ug/l	95
20) Methylene Chloride	2.794	84	10844	6.248	ug/l #	89
43) Isopropyl Acetate	6.354	43	17609	3.959	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
Data File : VX023746.D
Acq On : 16 Aug 2021 18:22
Operator : JC/MD
Sample : M3407-24
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
28S

Quant Time: Aug 17 06:19:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

