

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
 Data File : VX023747.D
 Acq On : 16 Aug 2021 18:45
 Operator : JC/MD
 Sample : M3407-27
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 29N

Quant Time: Aug 17 06:19:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	140233	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	240332	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229707	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88498	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	101887	58.245	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 116.500%	
35) Dibromofluoromethane	5.397	113	78812	52.448	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.900%	
50) Toluene-d8	8.653	98	302251	54.747	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 109.500%	
62) 4-Bromofluorobenzene	11.085	95	100026	53.282	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.560%	
Target Compounds						
						Qvalue
16) Acetone	2.392	43	14166	16.509	ug/l	# 86
20) Methylene Chloride	2.794	84	11985	7.180	ug/l	87
43) Isopropyl Acetate	6.354	43	23484	5.529	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
Data File : VX023747.D
Acq On : 16 Aug 2021 18:45
Operator : JC/MD
Sample : M3407-27
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
29N

Quant Time: Aug 17 06:19:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

