

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
 Data File : VX023748.D
 Acq On : 16 Aug 2021 19:09
 Operator : JC/MD
 Sample : M3407-30
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 29S

Quant Time: Aug 17 06:19:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	143236	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	250800	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	235952	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90888	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103732	58.056	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	116.120%
35) Dibromofluoromethane	5.397	113	80617	51.410	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.820%
50) Toluene-d8	8.653	98	311327	54.038	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.080%
62) 4-Bromofluorobenzene	11.085	95	103377	52.768	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.540%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	16756	19.118	ug/l	95
20) Methylene Chloride	2.794	84	11088	6.504	ug/l #	87
43) Isopropyl Acetate	6.355	43	21630	4.880	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
Data File : VX023748.D
Acq On : 16 Aug 2021 19:09
Operator : JC/MD
Sample : M3407-30
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
29S

Quant Time: Aug 17 06:19:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

