

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
 Data File : VX023749.D
 Acq On : 16 Aug 2021 19:32
 Operator : JC/MD
 Sample : M3408-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 31S

Quant Time: Aug 17 06:20:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	142523	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	244951	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	234990	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91853	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101526	57.106	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.220%
35) Dibromofluoromethane	5.397	113	79034	51.604	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.200%
50) Toluene-d8	8.653	98	309366	54.980	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	109.960%
62) 4-Bromofluorobenzene	11.085	95	101386	52.988	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.980%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	16254	18.638	ug/l	90
20) Methylene Chloride	2.794	84	11902	7.016	ug/l	93
43) Isopropyl Acetate	6.355	43	21600	4.989	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081621\
Data File : VX023749.D
Acq On : 16 Aug 2021 19:32
Operator : JC/MD
Sample : M3408-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
31S

Quant Time: Aug 17 06:20:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

