

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
 Data File : VX037007.D
 Acq On : 16 Aug 2023 10:55
 Operator : JC/MD
 Sample : VX0816WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0816WBL01

Quant Time: Aug 17 04:11:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	137958	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	252752	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	247765	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117772	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	96946	47.492	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.980%
35) Dibromofluoromethane	5.385	113	81231	46.758	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.520%
50) Toluene-d8	8.647	98	303061	49.192	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.380%
62) 4-Bromofluorobenzene	11.079	95	126108	54.255	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
Data File : VX037007.D
Acq On : 16 Aug 2023 10:55
Operator : JC/MD
Sample : VX0816WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0816WBL01

Quant Time: Aug 17 04:11:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

