

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
 Data File : VX037010.D
 Acq On : 16 Aug 2023 12:12
 Operator : JC/MD
 Sample : 04008-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P1632

Quant Time: Aug 16 12:45:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	125649	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	228908	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216294	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92740	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88836	47.783	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.560%
35) Dibromofluoromethane	5.385	113	73907	46.973	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.940%
50) Toluene-d8	8.647	98	275709	49.414	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.820%
62) 4-Bromofluorobenzene	11.079	95	106512	50.598	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.200%
Target Compounds						
16) Acetone	2.380	43	3723	4.826	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
Data File : VX037010.D
Acq On : 16 Aug 2023 12:12
Operator : JC/MD
Sample : 04008-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P1632

Quant Time: Aug 16 12:45:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

