

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
 Data File : VX037011.D
 Acq On : 16 Aug 2023 12:36
 Operator : JC/MD
 Sample : 04024-05 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE115D2-20230815

Quant Time: Aug 17 04:13:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	128498	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	235317	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221803	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96491	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90062	47.368	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.740%
35) Dibromofluoromethane	5.385	113	74345	45.965	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.920%
50) Toluene-d8	8.647	98	281284	49.040	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.080%
62) 4-Bromofluorobenzene	11.079	95	108905	50.325	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.660%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.325	101	10398	7.130	ug/l	99
12) 1,1-Dichloroethene	2.319	96	5374	3.735	ug/l	91
44) Trichloroethene	7.123	130	242890	132.481	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
Data File : VX037011.D
Acq On : 16 Aug 2023 12:36
Operator : JC/MD
Sample : 04024-05 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE115D2-20230815

Quant Time: Aug 17 04:13:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

