

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
 Data File : VX037013.D
 Acq On : 16 Aug 2023 13:24
 Operator : JC/MD
 Sample : 04024-08 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP05-20230815

Quant Time: Aug 17 04:14:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	127645	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	235969	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226361	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100017	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90632	47.987	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.980%
35) Dibromofluoromethane	5.385	113	74909	46.185	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.380%
50) Toluene-d8	8.647	98	282548	49.124	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.240%
62) 4-Bromofluorobenzene	11.079	95	111301	51.290	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.580%
Target Compounds						Qvalue
44) Trichloroethene	7.129	130	12996	7.069	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
Data File : VX037013.D
Acq On : 16 Aug 2023 13:24
Operator : JC/MD
Sample : 04024-08 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP05-20230815

Quant Time: Aug 17 04:14:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

