

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
 Data File : VX037016.D
 Acq On : 16 Aug 2023 14:35
 Operator : JC/MD
 Sample : 04024-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT15811-20230815

Quant Time: Aug 17 04:15:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	119830	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	221346	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	210630	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96777	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85555	48.253	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.500%
35) Dibromofluoromethane	5.385	113	71782	47.181	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.360%
50) Toluene-d8	8.647	98	265068	49.130	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.260%
62) 4-Bromofluorobenzene	11.079	95	106981	52.557	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.120%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1263	1.717	ug/l	94
19) Methyl tert-butyl Ether	3.117	73	1428	0.283	ug/l #	85
44) Trichloroethene	7.129	130	1632	0.946	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
Data File : VX037016.D
Acq On : 16 Aug 2023 14:35
Operator : JC/MD
Sample : 04024-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT15811-20230815

Quant Time: Aug 17 04:15:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

