

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
 Data File : VX037018.D
 Acq On : 16 Aug 2023 15:24
 Operator : JC/MD
 Sample : 04024-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE115D1-20230815

Quant Time: Aug 17 04:15:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	134464	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	247490	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	237671	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	112648	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92949	46.718	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.440%
35) Dibromofluoromethane	5.385	113	78478	46.133	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.260%
50) Toluene-d8	8.647	98	294927	48.890	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.780%
62) 4-Bromofluorobenzene	11.079	95	119015	52.292	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.580%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.325	101	10729	7.030	ug/l	99
12) 1,1-Dichloroethene	2.312	96	5344	3.549	ug/l	89
20) Methylene Chloride	2.788	84	1468	0.800	ug/l #	87
27) cis-1,2-Dichloroethene	4.489	96	2604	1.303	ug/l	87
30) Chloroform	5.099	83	3119	0.966	ug/l	89
44) Trichloroethene	7.129	130	286677	148.673	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
Data File : VX037018.D
Acq On : 16 Aug 2023 15:24
Operator : JC/MD
Sample : 04024-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE115D1-20230815

Quant Time: Aug 17 04:15:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

