

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
 Data File : VX037025.D
 Acq On : 16 Aug 2023 18:12
 Operator : JC/MD
 Sample : 04024-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT14911-20230815

Quant Time: Aug 17 04:18:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	130085	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	242935	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	237048	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	112347	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	91794	47.690	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.380%
35) Dibromofluoromethane	5.391	113	76596	45.871	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.740%
50) Toluene-d8	8.647	98	292545	49.404	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.800%
62) 4-Bromofluorobenzene	11.079	95	119585	53.528	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.060%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	918	1.149	ug/l #	85
19) Methyl tert-butyl Ether	3.117	73	2129	0.389	ug/l	91
20) Methylene Chloride	2.788	84	600	0.338	ug/l #	89
44) Trichloroethene	7.129	130	3817	2.017	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
Data File : VX037025.D
Acq On : 16 Aug 2023 18:12
Operator : JC/MD
Sample : 04024-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT14911-20230815

Quant Time: Aug 17 04:18:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

