

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
 Data File : VX037026.D
 Acq On : 16 Aug 2023 18:36
 Operator : JC/MD
 Sample : 04024-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW17811-20230815

Quant Time: Aug 17 04:18:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	127999	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	237964	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231571	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107125	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	91221	48.165	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.320%
35) Dibromofluoromethane	5.385	113	75983	46.455	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.900%
50) Toluene-d8	8.647	98	284066	48.974	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.940%
62) 4-Bromofluorobenzene	11.079	95	117606	53.742	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.480%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	2387	1.665	ug/l #	79
27) cis-1,2-Dichloroethene	4.489	96	55269	29.049	ug/l	82
44) Trichloroethene	7.129	130	565090	304.793	ug/l	98
45) 1,2-Dichloropropane	7.434	63	35739	18.872	ug/l	97
64) Tetrachloroethene	9.275	164	28746	16.643	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
Data File : VX037026.D
Acq On : 16 Aug 2023 18:36
Operator : JC/MD
Sample : 04024-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW17811-20230815

Quant Time: Aug 17 04:18:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

