

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
 Data File : VX037035.D
 Acq On : 17 Aug 2023 03:13
 Operator : JC/MD
 Sample : VX0816MBL02
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0816MBL02

Quant Time: Aug 17 05:08:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	146043	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	272951	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	266141	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	117910	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	103103	47.713	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.420%
35) Dibromofluoromethane	5.385	113	85656	45.656	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.320%
50) Toluene-d8	8.646	98	329652	49.548	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.100%
62) 4-Bromofluorobenzene	11.079	95	132227	52.678	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
Data File : VX037035.D
Acq On : 17 Aug 2023 03:13
Operator : JC/MD
Sample : VX0816MBL02
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0816MBL02

Quant Time: Aug 17 05:08:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

