

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
 Data File : VX037039.D
 Acq On : 17 Aug 2023 05:12
 Operator : JC/MD
 Sample : 03958-03 10X
 Misc : 6.30g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NEVINS-SB02-Z64-65

Quant Time: Aug 17 05:46:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	130469	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	241520	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	233675	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	108603	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	92761	48.051	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.100%
35) Dibromofluoromethane	5.385	113	76584	46.133	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.260%
50) Toluene-d8	8.647	98	291222	49.469	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.940%
62) 4-Bromofluorobenzene	11.079	95	116991	52.673	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.340%
Target Compounds						
					Qvalue	
67) Ethyl Benzene	10.195	91	40442	4.298	ug/l	98
69) o-Xylene	10.640	106	3674	1.048	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	18741	2.492	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
Data File : VX037039.D
Acq On : 17 Aug 2023 05:12
Operator : JC/MD
Sample : 03958-03 10X
Misc : 6.30g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NEVINS-SB02-Z64-65

Quant Time: Aug 17 05:46:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

