

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
 Data File : VX037042.D
 Acq On : 17 Aug 2023 06:24
 Operator : JC/MD
 Sample : 03958-06 10X
 Misc : 7.04g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NEVINS-SB02-Z21-23

Quant Time: Aug 17 07:13:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	135992	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	254696	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	248093	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	120603	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	97181	48.296	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.600%
35) Dibromofluoromethane	5.385	113	79733	45.545	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.100%
50) Toluene-d8	8.647	98	304766	49.091	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.180%
62) 4-Bromofluorobenzene	11.079	95	126505	54.010	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.020%
Target Compounds						
					Qvalue	
40) Benzene	6.038	78	60800	8.009	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
Data File : VX037042.D
Acq On : 17 Aug 2023 06:24
Operator : JC/MD
Sample : 03958-06 10X
Misc : 7.04g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NEVINS-SB02-Z21-23

Quant Time: Aug 17 07:13:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

