

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
 Data File : VX037045.D
 Acq On : 17 Aug 2023 07:35
 Operator : JC/MD
 Sample : 03975-04 10X
 Misc : 6.31g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NEVINS-SB03-Z82-83

Quant Time: Aug 17 08:01:34 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	137564	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	257110	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	259284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	126607	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	98172	48.231	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.460%
35) Dibromofluoromethane	5.385	113	80881	45.767	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.540%
50) Toluene-d8	8.647	98	310945	49.616	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.240%
62) 4-Bromofluorobenzene	11.079	95	133418	56.427	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.860%
Target Compounds						
					Qvalue	
40) Benzene	6.037	78	13007	1.697	ug/l	99
67) Ethyl Benzene	10.195	91	113790	10.898	ug/l	99
68) m/p-Xylenes	10.299	106	16649	4.274	ug/l	98
69) o-Xylene	10.640	106	10968	2.821	ug/l	90
73) Isopropylbenzene	10.963	105	24117	2.302	ug/l	98
78) n-propylbenzene	11.305	91	14022	1.157	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	24588	2.805	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	78996	9.011	ug/l	96
86) p-Isopropyltoluene	12.006	119	11311	1.280	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
Data File : VX037045.D
Acq On : 17 Aug 2023 07:35
Operator : JC/MD
Sample : 03975-04 10X
Misc : 6.31g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NEVINS-SB03-Z82-83

Quant Time: Aug 17 08:01:34 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

