

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
 Data File : VX037053.D
 Acq On : 17 Aug 2023 10:47
 Operator : JC/MD
 Sample : 03958-07
 Misc : 7.94g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NEVINS-SB02-Z10-12

Quant Time: Aug 17 12:23:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	126386	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	229350	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	233375	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	130347	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	81957	43.826	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	87.660%		
35) Dibromofluoromethane	5.385	113	68821	43.656	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	87.320%		
50) Toluene-d8	8.647	98	277698	49.674	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.340%		
62) 4-Bromofluorobenzene	11.085	95	128782	61.059	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	122.120%		
Target Compounds						
					Qvalue	
39) Methylcyclohexane	7.373	83	7090	2.427	ug/l #	92
40) Benzene	6.031	78	278553	40.747	ug/l	98
52) Toluene	8.720	92	7559	1.765	ug/l	98
67) Ethyl Benzene	10.195	91	1439613	153.190	ug/l	98
68) m/p-Xylenes	10.299	106	196232	55.965	ug/l	98
69) o-Xylene	10.640	106	134199	38.343	ug/l	99
73) Isopropylbenzene	10.963	105	214532	19.887	ug/l	98
78) n-propylbenzene	11.305	91	115497	9.258	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	198906	22.041	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	848586	94.024	ug/l	95
86) p-Isopropyltoluene	12.012	119	107998	11.871	ug/l	95
89) n-Butylbenzene	12.317	91	171584	21.069	ug/l #	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
Data File : VX037053.D
Acq On : 17 Aug 2023 10:47
Operator : JC/MD
Sample : 03958-07
Misc : 7.94g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 56 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NEVINS-SB02-Z10-12

Quant Time: Aug 17 12:23:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

