

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
 Data File : VX037005.D
 Acq On : 16 Aug 2023 10:03
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 17 04:04:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	70	0.00
2 T	Dichlorodifluoromethane	0.661	0.648	2.0	72	0.00
3 P	Chloromethane	0.601	0.520	13.5	64	0.00
4 C	Vinyl Chloride	0.612	0.588	3.9#	68	0.00
5 T	Bromomethane	0.365	0.418	-14.5	83	-0.02
6 T	Chloroethane	0.297	0.390	-31.3#	88	-0.01
7 T	Trichlorofluoromethane	0.945	1.092	-15.6	82	0.00
8 T	Diethyl Ether	0.365	0.435	-19.2	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.567	0.627	-10.6	78	0.00
10 T	Methyl Iodide	0.589	0.767	-30.2#	86	0.00
11 T	Tert butyl alcohol	0.175	0.186	-6.3	75	-0.02
12 CM	1,1-Dichloroethene	0.560	0.575	-2.7#	74	0.00
13 T	Acrolein	0.100	0.106	-6.0	78	0.00
14 T	Allyl chloride	1.102	0.976	11.4	64	0.00
15 T	Acrylonitrile	0.352	0.382	-8.5	75	0.00
16 T	Acetone	0.307	0.356	-16.0	84	-0.01
17 T	Carbon Disulfide	1.545	1.296	16.1	60	0.00
18 T	Methyl Acetate	1.291	1.364	-5.7	74	0.00
19 T	Methyl tert-butyl Ether	2.104	2.502	-18.9	83	0.00
20 T	Methylene Chloride	0.683	0.699	-2.3	76	0.00
21 T	trans-1,2-Dichloroethene	0.626	0.666	-6.4	77	0.00
22 T	Diisopropyl ether	2.093	2.096	-0.1	71	0.00
23 T	Vinyl Acetate	1.606	1.650	-2.7	71	0.00
24 P	1,1-Dichloroethane	1.174	1.236	-5.3	74	0.00
25 T	2-Butanone	0.498	0.538	-8.0	75	-0.01
26 T	2,2-Dichloropropane	1.106	1.167	-5.5	75	0.00
27 T	cis-1,2-Dichloroethene	0.743	0.827	-11.3	80	0.00
28 T	Bromochloromethane	0.568	0.510	10.2	63	-0.01
29 T	Tetrahydrofuran	0.321	0.330	-2.8	71	0.00
30 C	Chloroform	1.201	1.408	-17.2#	84	0.00
31 T	Cyclohexane	1.024	0.924	9.8	64	0.00
32 T	1,1,1-Trichloroethane	1.100	1.234	-12.2	80	0.00
33 S	1,2-Dichloroethane-d4	0.740	0.727	1.8	71	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	69	0.00
35 S	Dibromofluoromethane	0.344	0.350	-1.7	72	0.00
36 T	1,1-Dichloropropene	0.512	0.543	-6.1	75	0.00
37 T	Ethyl Acetate	0.580	0.595	-2.6	70	0.00
38 T	Carbon Tetrachloride	0.545	0.624	-14.5	80	0.00
39 T	Methylcyclohexane	0.637	0.575	9.7	63	0.00
40 TM	Benzene	1.490	1.610	-8.1	75	0.00
41 T	Methacrylonitrile	0.310	0.335	-8.1	73	0.00
42 TM	1,2-Dichloroethane	0.545	0.675	-23.9	85	0.00
43 T	Isopropyl Acetate	0.998	1.019	-2.1	71	0.00
44 TM	Trichloroethene	0.390	0.446	-14.4	80	0.00
45 C	1,2-Dichloropropane	0.398	0.422	-6.0#	74	0.00
46 T	Dibromomethane	0.278	0.335	-20.5	83	0.00
47 T	Bromodichloromethane	0.572	0.664	-16.1	80	0.00
48 T	Methyl methacrylate	0.464	0.488	-5.2	72	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
 Data File : VX037005.D
 Acq On : 16 Aug 2023 10:03
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 17 04:04:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.011	-22.2	79	-0.04
50 S	Toluene-d8	1.219	1.096	10.1	62	0.00
51 T	4-Methyl-2-Pentanone	0.552	0.596	-8.0	72	0.00
52 CM	Toluene	0.934	1.054	-12.8#	77	0.00
53 T	t-1,3-Dichloropropene	0.646	0.737	-14.1	77	0.00
54 T	cis-1,3-Dichloropropene	0.682	0.761	-11.6	77	0.00
55 T	1,1,2-Trichloroethane	0.393	0.465	-18.3	82	0.00
56 T	Ethyl methacrylate	0.639	0.733	-14.7	77	0.00
57 T	1,3-Dichloropropane	0.654	0.770	-17.7	80	0.00
58 T	2-Chloroethyl Vinyl ether	0.315	0.341	-8.3	74	0.00
59 T	2-Hexanone	0.410	0.452	-10.2	74	0.00
60 T	Dibromochloromethane	0.431	0.530	-23.0	83	0.00
61 T	1,2-Dibromoethane	0.407	0.506	-24.3	85	0.00
62 S	4-Bromofluorobenzene	0.460	0.471	-2.4	68	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	68	0.00
64 T	Tetrachloroethene	0.373	0.439	-17.7	81	0.00
65 PM	Chlorobenzene	1.102	1.266	-14.9	79	0.00
66 T	1,1,1,2-Tetrachloroethane	0.426	0.519	-21.8	84	0.00
67 C	Ethyl Benzene	2.013	2.238	-11.2#	76	0.00
68 T	m/p-Xylenes	0.751	0.862	-14.8	78	0.00
69 T	o-Xylene	0.750	0.858	-14.4	77	0.00
70 T	Styrene	1.237	1.449	-17.1	78	0.00
71 P	Bromoform	0.345	0.417	-20.9	80	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	73	0.00
73 T	Isopropylbenzene	4.138	4.196	-1.4	76	0.00
74 T	N-amyl acetate	1.930	1.768	8.4	68	0.00
75 P	1,1,2,2-Tetrachloroethane	1.398	1.468	-5.0	80	0.00
76 T	1,2,3-Trichloropropane	1.176	1.295	-10.1	82	0.00
77 T	Bromobenzene	0.942	1.055	-12.0	82	0.00
78 T	n-propylbenzene	4.786	4.740	1.0	73	0.00
79 T	2-Chlorotoluene	2.900	2.991	-3.1	78	0.00
80 T	1,3,5-Trimethylbenzene	3.462	3.574	-3.2	76	0.00
81 T	trans-1,4-Dichloro-2-butene	0.519	0.499	3.9	70	0.00
82 T	4-Chlorotoluene	3.304	3.490	-5.6	79	0.00
83 T	tert-Butylbenzene	3.401	3.394	0.2	75	0.00
84 T	1,2,4-Trimethylbenzene	3.462	3.600	-4.0	77	0.00
85 T	sec-Butylbenzene	4.241	4.063	4.2	70	0.00
86 T	p-Isopropyltoluene	3.490	3.477	0.4	72	0.00
87 T	1,3-Dichlorobenzene	1.745	1.944	-11.4	82	0.00
88 T	1,4-Dichlorobenzene	1.732	1.965	-13.5	83	0.00
89 T	n-Butylbenzene	3.124	2.955	5.4	68	0.00
90 T	Hexachloroethane	0.707	0.610	13.7	64	0.00
91 T	1,2-Dichlorobenzene	1.684	1.940	-15.2	85	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.325	0.350	-7.7	82	0.00
93 T	1,2,4-Trichlorobenzene	1.038	1.158	-11.6	81	0.00
94 T	Hexachlorobutadiene	0.429	0.448	-4.4	79	0.00
95 T	Naphthalene	3.616	4.390	-21.4	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081623\
Data File : VX037005.D
Acq On : 16 Aug 2023 10:03
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 17 04:04:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.023	1.163	-13.7	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6