

Data File : VX017939.D
Acq On : 17 Aug 2020 15:23
Operator : JC/SP
Sample : L3701-08
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JLTX7

Quant Time: Aug 18 01:41:45 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR080720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Aug 18 00:59:21 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	270589	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	272796	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	116228	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	83096	4.44	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.80%
7) Chloroethane-d5	1.70	69	72337	4.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.40%
11) 1,1-Dichloroethene-d2	2.34	63	128899	3.35	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.00%
20) 2-Butanone-d5	4.54	46	281651	56.72	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.44%
24) Chloroform-d	5.14	84	195122	5.14	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.80%
26) 1,2-Dichloroethane-d4	6.03	65	114203	5.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.60%
32) Benzene-d6	6.05	84	360473	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.20%
36) 1,2-Dichloropropane-d6	7.37	67	120746	5.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.40%
41) Toluene-d8	8.70	98	309261	4.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.80%
45) trans-1,3-Dichloropropene-	8.99	79	46168	4.61	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.20%
48) 2-Hexanone-d5	9.43	63	221638	53.19	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.38%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	92087	5.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.60%
65) 1,2-Dichlorobenzene-d4	12.36	152	127156	5.68	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.60%

Target Compounds					Qvalue
25) Chloroform	5.18	83	10306	0.228 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX017939.D
Acq On : 17 Aug 2020 15:23
Operator : JC/SP
Sample : L3701-08
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JLTX7

Quant Time: Aug 18 01:41:45 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR080720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Aug 18 00:59:21 2020
Response via : Initial Calibration

