

Data File : VX017945.D
Acq On : 17 Aug 2020 18:16
Operator : JC/SP
Sample : L3701-14
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JLTY3

Quant Time: Aug 18 02:11:02 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR080720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Aug 18 00:59:21 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	247774	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	252987	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	109195	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	78464	4.58	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.60%
7) Chloroethane-d5	1.69	69	68702	5.05	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	101.00%
11) 1,1-Dichloroethene-d2	2.34	63	118580	3.36	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.20%
20) 2-Butanone-d5	4.53	46	308005	67.74	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	135.48%#
24) Chloroform-d	5.14	84	189419	5.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.00%
26) 1,2-Dichloroethane-d4	6.03	65	112095	5.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	117.40%
32) Benzene-d6	6.05	84	340071	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
36) 1,2-Dichloropropane-d6	7.37	67	116921	5.45	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.00%
41) Toluene-d8	8.70	98	290007	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.80%
45) trans-1,3-Dichloropropene-	8.99	79	45698	4.92	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.40%
48) 2-Hexanone-d5	9.43	63	242085	62.65	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	125.30%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	95781	5.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	117.40%
65) 1,2-Dichlorobenzene-d4	12.36	152	133174	6.34	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	126.80%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX017945.D
Acq On : 17 Aug 2020 18:16
Operator : JC/SP
Sample : L3701-14
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JLTY3

Quant Time: Aug 18 02:11:02 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR080720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Aug 18 00:59:21 2020
Response via : Initial Calibration

